

장비 사양서

장비심의

제2020-임9호 / 혈액가스분석기 1대(Blood Gas Analyzer)

1. 기 본 사 양 (기본적구성)

- 1) Measured Parameters : Ph, PO₂, PCO₂, NA⁺, K⁺, CA⁺⁺, CL⁻, Glucose, Lac, Co-Oxy
- 2) Sample Size : 175ul or 195ul
- 3) Sample Type : Whole blood
- 4) Reagent Type : 3개 or 5개
- 5) Calibration Interval : 1-point 0.5hour/1, full-point 8hour or 1-point 4hour, 2-point 8hour
- 6) Calibration System Type : 1-point cal, 2-point cal
- 7) Measurement Time : 60sec or 135sec
- 8) QC : Auto QC or half Auto QC
- 9) Sampling Type : Syringe & Capillary type, Aspiration type
- 10) Electrode Type : Ready sensor or Membrane sensor
- 11) Display : Color Touch Screen
- 12) Maintence : 가능
- 13) Clot Detection : Probe Detection & Clearence or Warning
- 14) Data : Patient 5,000data, QC 2,100data or Patient 2,000data, QC 1,500data
- 15) Co-ox Upgrade : 가능
- 16) Inter face System : DMS, LIS, HIS

2. 구 성 품

- | | |
|-------------------------|----------------|
| 1) Analyzer Unit | · · · · · 1Set |
| 2) Standard Accessories | · · · · · 1Set |

3. 특 이 사 항

- 1) Operating / Service Manual 각각 제공한다.
- 2) 모든 장비의 구성 및 Accessary는 사양대로 제공할 것.
- 3) Warranty는 임대기간 무상수리.

장비 사양서

장비심의

제2020-임9호 / 혈액가스분석기 1대(Blood Gas Analyzer)

- 4) 회사의 최신장비, 최신 Model을 납품.
- 5) PACS or EMR Interface 연동.
- 6) 사양 검토 시 우리병원에 제출한 모든 것.
- 7) All Software Install program & All service key 제공.
- 8) 소모품 및 시약 물품내역서 첨부.

2020. 12.