

장비 사양서

장비심의

제2020-임6호 / 내시경 세척기(Endoscopic Washer)

1. 기 본 사 양 (기본적구성)

1) Cleaning & Disinfection Method :

Outer surface = Ultrasonic + Running fluid cleaning

Inside channel = Fluid flushing cleaning

Valves = Ultrasonic + fluid flushing cleaning

2) Cleaning / Disinfection time : Disinfection + 10 min

3) Disinfection time setting : 5 - 60 minutes

4) Disinfection inside system : Automatic

5) Supply water flow : 17 ℓ /min

6) Supply water pressure : Between 0.1 MPa and 0.5MPa

7) Cleaning tub capacity : 14 liters

8) Disinfection solution tank capacity : 17.5 liters

9) Using Disinfectant : Acecide 875ml (전용소독기)

10) Alcohol flushing : Automatic

11) Visual leakage detection : Available

12) Scope compatibility : Yes

13) Dimensions (W×D×H mm) : 450×765×977

2. 구 성 품

1) Endoscopic Washer 1Set

2) Standard Accessories 1Set

3. 특 이 사 항

- 1) Operating / Service Manual 각각 제공한다.
- 2) 모든 장비의 구성 및 Accessary는 사양대로 제공할 것.
- 3) Warranty는 임대기간 무상수리.
- 4) 회사의 최신장비, 최신 Model을 납품.
- 5) PACS or EMR Interface 연동.

장비 사양서

장비심의

제2020-임6호 / 내시경 세척기(Endoscopic Washer)

- 6) 사양 검토 시 우리병원에 제출한 모든 것.
- 7) All Software Install program & All service key 제공.
- 8) 소모품 및 시약 물품내역서 첨부.

2020. 12.